

## VERIFICATION OF COMPLIANCE

No.: LVD GZES2012035917HS  
Applicant: Guangdong UBORUI Technology Co., Ltd.  
6 Floor, 3#, 1 Haiyi Street, Lanhe Town, Nansha, Guangzhou,  
Guangdong, China  
Manufacturer: Guangdong UBORUI Technology Co., Ltd.  
6 Floor, 3#, 1 Haiyi Street, Lanhe Town, Nansha, Guangzhou,  
Guangdong, China  
Product Name: Hydrogen Inhaler  
Product Description: Electrolyser  
Model No.: H300S, H300D  
Trade Mark: **YBR**  
Rating: 220 V – 240 V; 50 Hz / 60 Hz; 140 W;  
Protection against Electric Shock: Class II  
Additional Information: --  
Sufficient samples of the product have been tested and found to be in conformity with  
Test Standard: EN 60335-1: 2012 + A11: 2014 + A13: 2017 + A1: 2019 + A14:  
2019 + A2: 2019  
EN 60335-2-108: 2008  
EN 62233: 2008

as shown in the

Test Report Number(s): GZES201203591701

This Verification of Compliance has been granted to the applicant based on the results of tests, performed by Laboratory of SGS-CSTC Standards Technical Services Co., Ltd. on sample of the above-mentioned product in accordance with the provisions of the relevant harmonized standards under the Low Voltage Directive 2014/35/EU. The CE marking as shown below can be affixed, under the responsibility of the manufacturer, after completion of an EC Declaration of Conformity and compliance with all relevant EC Directives. The affixing of the CE marking presumes in addition that the conditions in annexes III and IV of the Directive are fulfilled.



David Guo  
Senior Technical Manager  
SGS-CSTC



2021-06-25

Copyright of this verification is owned by SGS-CSTC Standards Technical Services Co., Ltd. and may not be reproduced other than in full and with the prior approval of the General Manager. This verification is subjected to the governance of the General Conditions of Services, printed overleaf.

Member of SGS Group (Société Générale de Surveillance)